

## Certificate

## High Containment Biosafety Level 3 Virology BSL-3 Enhanced

This is to certify that the biosafety level 3 virology laboratory, designed and built by ERLA Technologies Sdn. Bhd. according to ANZS 2243.3:2002, operated by the

National Public Health Laboratory
Makmal Kesihatan Awam Kebangsaan
Kementerian Kesihatan Malaysia
Lot 1853 Kg. Melayu Sg. Buloh
47000 Sg. Buloh, Selangor

complies with the high containment BSL-3 laboratory standards set by the WHO Guidelines (2004) and the US NIH Biosafety Level 3-Laboratory Certification Requirements for their own laboratories, and in part exceeds these requirements.

The architecture, the facility, the engineering and the implementation as a whole is in compliance with international state-of-the-art biosafety principles. The biosafety management system was not part of the scope of the certification.

The certification inspection was performed on: 7 to 10 August 2007 and 7 September 2007.

The certificate is valid until 6 September 2008.

Basler & Hofmann Singapore Pte Ltd, MOH-Approved External Certifiers

Dr. Felix K. Gmünder

Approved Certifier and Biosafety Professional

mude